




# Medical Device Security

Mark Olson  
hospital CISO



## General System Counts

Systems with AV.....6398  
 Printers.....2074  
 Medical equipment...**905**  
 Misc.....2460  
 -----  
 Total Devices:.....11837

## OS Makeup – Medical

Windows 95.....1  
 Windows 98 .....15  
 Windows 2000.....23  
 Windows CE.....9  
 Windows Vista.....0  
 Windows XP.....600  
 Windows XP SP1.....0  
 Windows XP SP2....15  
 Windows XP SP3....1  
 -----  
 Total..... 664

## Average Time to Infection

Clinical Systems , 510K, no AV.: **12 days**  
 Systems running AV/Patches.....: **300+ days**

**Ideally:** FDA 510K is updated to include a requirement for the provision of industry accepted security controls for devices utilizing embedded operating systems or other controllers associated with a medical device

**Alternatively:** The FDA issues a clear statement to the community that FDA 510K is not jeopardized by permitting Anti-Virus or Operating System patching to the supporting systems associated with a certified medical device